

MATERIAL SAFETY DATA SHEET

PDI - PROFESSIONAL DISPOSABLES, INC
THE HEALTHCARE DIVISION OF
NICE-PAK PRODUCTS, INC
TWO NICE-PAK PARK
ORANGEBURG NY 10962-1376
845-365-1700

REVISION DATE: 4/02

I. PRODUCT IDENTIFICATION

PRODUCT/TRADE NAME: CLEANSING TOWELETTE D31000

HAZARD RATING (NFPA)

HEALTH: 1

FLAMMABILITY: 2

REACTIVITY: 0

SPECIFIC: NONE

EMERGENCY OR INFORMATION TELEPHONE NO: 845-365-1700 (M-F DAYTIME) AT OTHER TIMES, CONTACT
THE LOCAL POISON CONTROL CENTER

CHEMICAL NAME: PROPRIETARY MIXTURE

II HAZARDOUS INGREDIENTS PER 29 CFR 1910.1200

Hazardous Ingredients	%	ACGIH TLV	CAS NUMBER
SD ALCOHOL 40	5	1000 ppm	64-17-5

III-PHYSICAL/CHEMICAL CHARACTERISTICS

COLOR/ODOR/APPEARANCE: TOWELETTE SATURATED WITH A HAZY LIQUID WITH THYMOL/CAMPHOR ODOR
BOILING POINT: N/A

FLASH POINT: 169⁰ F

VAPOR DENSITY N/A

EVAPORATION RATE: N/A

SOLUBILITY IN WATER: COMPLETE

SPECIFIC GRAVITY (H₂O = 1): 0.994

IV FIRE & EXPLOSION HAZARD DATA

FLASH POINT(Method Used): 169⁰ F TAG CLOSED CUP ; LEL: N/A UEL: N/A

EXTINGUISHING MEDIA: DRY CHEMICAL OR ALCOHOL TYPE FOAM, CARBON DIOXIDE

SPECIAL FIRE FIGHTING PROCEDURES: NONE

UNUSUAL FIRE AND EXPLOSION HAZARDS: NONE

V-REACTIVITY DATA

STABILITY: STABLE

CONDITIONS TO AVOID : NONE KNOWN

INCOMPATIBILITY: NONE KNOWN

HAZARDOUS DECOMPOSITION OR BYPRODUCTS: NONE KNOWN

POLYMERIZATION: WILL NOT OCCUR

CONDITIONS TO AVOID: NONE KNOWN

CLEANSING TOWELETTE

VI-HEALTH HAZARD DATA

EFFECTS OF OVER EXPOSURE:

SKIN : IF RASH APPEARS, DISCONTINUE USE

EYES : WILL CAUSE EYE STING IF SPLASHED

INHALATION: NONE

INGESTION: NOT NORMAL ROUTE OF ENTRY

EMERGENCY AND FIRST AID PROCEDURES:

SKIN CONTACT: IF RASH OR IRRITATION DEVELOPS, DISCONTINUE USE

EYE CONTACT: FLUSH WITH COLD WATER FOR 15 MINUTES

INHALATION: MOVE TO FRESH AIR

INGESTION: IF SWALLOWED CALL POISON CONTROL CENTER

TARGET ORGANS; N/A

KNOWN CARCINOGEN: NO

VII-SPILL AND DISPOSAL PROCEDURE

SPILL CONTROL: ELIMINATE ALL SOURCES OF IGNITION

WASTE DISPOSAL METHOD: FLUSH SPILLS WITH WATER. HANDLE AS FLAMMABLE LIQUID. FOLLOW LOCAL STATE AND FEDERAL REGULATIONS.

HANDLING AND STORAGE: STORE AWAY FROM HEAT AND IGNITION SOURCES

VIII-CONTROL MEASURES/PROTECTION

RESPIRATION: USE IF PERMISSIBLE EXPOSURE LEVEL IS EXCEEDED WHEN HANDLING BULK LIQUID.

VENTILATION: RECOMMENDED

PROTECTIVE GLOVES: NONE REQUIRED

EYE PROTECTION: ONLY IF SPLASHING IS EXPECTED

HYGIENIC PRACTICES: GOOD HOUSEKEEPING PRACTICES SHOULD BE FOLLOWED

OTHER: DON'T ALLOW LARGE QUANTITIES OF WASTE TO ACCUMULATE

IX TRANSPORT/SHIPPING

DOT SHIPPING NAME: CONSUMER COMMODITY

TECHNICAL SHIPPING NAME: N/A

DOT SHIPPING CLASSIFICATION: DOT NOT REGULATED

DOT ID NO.:N/A

DOT LABEL REQUIREMENTS: N/A

UN/NA NUMBER REGULATIONS: N/A

REPORTABLE QUANTITY: N/A

X DISCLAIMER

THE INFORMATION FURNISHED HEREIN IS BELIEVED TO BE ACCURATE AND REPRESENTS THE BEST DATA CURRENTLY AVAILABLE TO US. NO WARRANTY, EXPRESSED OR IMPLIED IS MADE AND NICE-PAK PRODUCTS, INC. ASSUMES NO LEGAL RESPONSIBILITY OR LIABILITY RESULTING FROM ITS USE.

Section V — Reactivity Data

Stability	Unstable		Conditions to Avoid
	Stable	X	

Incompatibility (Materials to Avoid)

Hazardous Decomposition or Byproducts

Hazardous Polymerization	May Occur		Conditions to Avoid
	Will Not Occur	X	

Section VI — Health Hazard Data

Route(s) of Entry: Inhalation? TOPICALLY APPLIED Skin? Ingestion?

Health Hazards (Acute and Chronic) NONE

Carcinogenicity: N/A NTP? IARC Monographs? OSHA Regulated?

Signs and Symptoms of Exposure NONE

Medical Conditions Generally Aggravated by Exposure NONE

Emergency and First Aid Procedures NONE

Section VII — Precautions for Safe Handling and Use

Steps to Be Taken in Case Material Is Released or Spilled CLEAN UP WITH WATER

Waste Disposal Method ORDINARY TRASH DISPOSAL OR AS APPROPRIATE PER LOCAL REGULATIONS

Precautions to Be Taken in Handling and Storing NONE

Other Precautions

Section VIII — Control Measures

Respiratory Protection (Specify Type) N/A

Ventilation	Local Exhaust	Special
N/A	Mechanical (General)	Other

Protective Gloves N/A Eye Protection N/A

Other Protective Clothing or Equipment N/A

Work/Hygienic Practices N/A


Bard® Foley Catheter Inflation/Deflation Guidelines

Proper Catheter Inflation

Ensure that the Bard Foley catheter balloon is positioned well within the patient's bladder. Slowly, with a gentle, constant force, inflate the Bard Foley catheter balloon with the volume prescribed on the package. **Note that a 5cc balloon must be inflated with between 9cc and 10cc of sterile water.** Improperly inflated Bard Foley catheter balloons may cause drainage and deflation difficulties.


Balloon Size	Recommended Inflation Capacities
3cc balloon	5cc sterile water
5cc balloon	10cc sterile water
30cc balloon	35cc sterile water
75cc balloon	80cc sterile water

Properly Inflated



5cc Foley catheter
inflated with
10cc of water

Under Inflated



5cc Foley catheter
inflated with
only 5cc of water



C. R. Bard, Inc.
Covington, GA 30014
1-800-526-4455
www.bardmedical.com

Please see reverse S

Catheter Deflation

- Select a luer slip syringe.
- Slide the plunger of the syringe up and down the barrel of the syringe several times to "loosen it up."
- Compress the plunger all the way and then pull back the plunger slightly so that it does not adhere to the front of the syringe barrel.
- Gently insert the syringe in the catheter valve.
- Do not use more force than is required to make the syringe "stick" in the valve.
- Allow the pressure within the balloon to force the plunger back and fill the syringe with water.
- If you notice slow or no deflation, re-seat the syringe gently. Once again, allow the balloon to deflate slowly on its own.
- If the balloon does not deflate, reposition the patient.
- Ensure that the catheter is not in traction—the proximal end of the catheter is not compressed within the bladder neck.
- Ensure that urine flows freely.
- Attempt to deflate the balloon by using the pressure in the balloon to force water into the syringe as described above.
- If the balloon still fails to deflate, apply very gentle, slow aspiration. Aspiration that is too rapid, or too forceful, may cause the inflation lumen within the Foley catheter to collapse.
- If permitted by hospital protocol, the valve arm may be severed.
- If this fails, contact an adequately trained professional for assistance, as directed by hospital protocol.



C. R. Bard, Inc.
Covington, GA 30014
1-800-526-4455
www.bardmedical.com

Please consult product labels and inserts for any indications, contraindications, warnings, cautions and directions for use.

Bard is a registered trademark of C. R. Bard, Inc. or an affiliate.
© 1999 C. R. Bard, Inc. (All rights reserved). Printed in U.S.