

Coat end of cone with lubricant before inserting into stoma. If resistance is met, discontinue irrigation and consult your physician.